

How to determine if the customer's existing controller will be compatible with the new summing card HI6020IT-SC1 if they are trying to replace their existing HI215IT-SC1. Factory is not always able to look up the serial numbers so this would be the best determination.

I wasn't able to find out anything by that serial number. The customer can find out what firmware level they have in the unit by going into the self- test menu.

1. Press the (-) key or Test/Clr button
2. Display will show SLFTST
3. Press Enter
4. Display will automatically scroll through its settings, you are looking for the screen that says Ver 1.XX
5. This is your firmware revision. Your unit has to have at least firmware revision 1.46 or higher for the IT to function properly.
6. Press exit key to exit self-test mode.

POB 1778 Greer, SC 29652  
414 West Poinsett Street Greer, SC 29650

[mike@traskinst.com](mailto:mike@traskinst.com)

[www.traskinstrumentation.com](http://www.traskinstrumentation.com)

Ph: 864-848-3993

864-848-9569

Fax: